

August 2022

Important update on the expiry of EMERADE™ Epinephrine Autoinjector

In July 2022, EMERADE™ 0.3 mg (DIN 02458446) and 0.5 mg (DIN 02458454) autoinjectors received an authorization from Health Canada to extend the expiration of the product from 18 to 24 months.

To avoid a drug shortage, Bausch Health Canada Inc. has extended the expiration dates by 6 months on EMERADE™ 0.3 mg and 0.5 mg on five (5) lots currently in the market (please refer to the table below). Health Canada has been notified and has no objection with this proposition.

This means that if your EMERADE™ Epinephrine Autoinjector is from any of the lots stated in the table below, a new expiration date applies. Refer to “New Expiration date” Column. Note that EMERADE™ is safe to use up until the new expiration date.

	Lot Number	Original Expiration Date (on label)	New Expiration date
EMERADE 0.3 MG (DIN 02458446)	Z0034B2C	June 2023	Dec 2023
	Z0076B1A	August 2023	Feb 2024
EMERADE 0.5 MG (DIN 02458454)	Z0034C1A	June 2023	Dec 2023
	Z0076C2A & Z0088C3A	August 2023	Feb 2024

*Note: expiration dates are valid through the last day of the month listed

Where do I find the lot number of my EMERADE™ Epinephrine Autoinjector?

The lot numbers are printed on the device, below the original expiration date.

If you have additional questions, you can always reach us by email at canada.customerservice@bauschhealth.com or by phone: 1-800-361-4261.

Sincerely,



Tibor Kapusy
Director, Regulatory Affairs



Mark Parent
Marketing Director, Specialty